



WINDENSE

Windense International

306 Port Royal Drive
Matthews, NC 28105
Tel (704) 708-5117
Fax (419) 831-2370

Updating Correction Factors

Background

Windense and Scandense store correction factors for Kodak process control tools. When installation files and CDs are created, they include the then-current control tool batch correction factors. Since Kodak is introducing correction factors regularly, the data in Windense and Scandense is soon no longer current.

If you are networked to a central location by FTP, Email or LAN, you may already be having correction factors updated automatically. If not, you can update them manually according to the instructions below.

Updating Correction Factors Manually

Download the current correction factors (win_cfs.zip) from www.windenseint.com

Scandense –

- Shutdown Scandense and Windense
- Copy win_cfs.zip to the \Scandens\Temp directory
- Launch Scandense

Windense

- Shutdown Windense
- Copy win_cfs.zip to the \Windense\Remin directory
- Launch Windense

RHSpies
WindenseNA@Yahoo.com

Corporate Headquarters:
19 Grenadier Park, Grenville Gate, Glasgow G72 8EP, Scotland
Web site: www.Windenseint.com - e-mail: WindenseNA@Yahoo.com