

Windense - [#1 phntm US (auto) 06/30/00 09:26]

File Tools Windows Help

Phantom Image Results phntm US (auto)

General Hospital GH #1 06/30/00 09:26

Exposure Factors
 kVp : 26.00 mAs : 126.00
 Photocell Position : Center

Density Information
 Disk Background (A) : 1.65 Disc Density (B) : 1.17 Density Difference (A-B) : 0.48
 Background Density : 1.65 AEC : 0

Summary Information
 Total test objects : **12.00** Technique : Non-Grid User No. : 1

Region Materials	Count
1.56 mm nylon fiber	1.0
1.12 mm nylon fiber	1.0
0.89 mm nylon fiber	1.0
0.75 mm nylon fiber	1.0
0.54 mm nylon fiber	0.5
0.40 mm nylon fiber	0.0
0.54 mm AL2O3 specks	1.0
0.40 mm AL2O3 specks	1.0
0.32 mm AL2O3 specks	1.0
0.24 mm AL2O3 specks	0.5
0.16 mm AL2O3 specks	0.0
2.00 mm mass	1.0
1.00 mm mass	1.0
0.75 mm mass	1.0
0.50 mm mass	1.0
0.25 mm mass	0.0

	Difference from reference	Readings	Reference
Den Diff	+0.00	+0.48	+0.48
Back Den	-0.03	+1.65	+1.68

521

Taskbar: C:\M\Installer\Me... C:\Users\Bob\Des... Healthcare - Intern... Inbox - Microsoft ... Windense - [#1 ph... WindenseHealthC... Adobe Acrobat Sta... 2:47 PM

The above is a US Phantom. Windense automatically presents the appropriate phantom format depending on your country.